

# **ANALYSIS CERTIFICATE**

**CODE: 181693 BATCH:** 0001365644

PRODUCT: Sodium Hydroxide 0.1 mol/l (0.1N) volumetric solution

ISSUE DATE: 18/04/2018 RETEST DATE: 04/2024

SPECIFICATIONS Titer at 20°C	GUARANTEE VALUE	ACTUAL VALUE 1,001
Uncertainty	± 0,0015	
Determination method	E	

#### Balance

Calibrated periodically against traceable weights.

# Volumetric glassware

Class A according NBS, document NBSIR-74661-Dec 1974

#### **Determination of concentration**

Assay by Hydrochloric Acid titration checked itself with Tris-Hydroxymethyl) Aminomethane standard.

traceable with primary standard NIST, SRM: 723e

Indicator: Bromophenol blue

DETERMINATION METHODS	SUBSTANCE	SRM/NIST	
A Argentimetric titration	Potassium Chloride	999b	
B Permanganometric titration	Sodium Oxalate	8040	
I Redox titration	Potassium Dichromate	136f	
C Complexometric titration	Zinc	682	
D Iodometric titration	Arsenic III Oxide	3103a	
E Acidimetry	Tris (Hydroxymethyl) Aminomethane	723e	
H Alkalimetry	Potassium Hydrogen Phthalate	841	
F Precipitation titrimetry	Potassium Chloride	999b	
G Non aqueous acid titration	Potassium Hydrogen Phthalate	841	
J Non aqueous alkali titration	Benzoic Acid	350b	
K Acidimetry	Sodium Carbonate	351a	

### Notes

 $\hfill \square$  All chemicals used for preparing this solution are for analysis, ACS quality.

 $\ \square$  The stability of prepared solution must be checked periodically by the user.

Don't use the solution if it is contaminated accidentally.

Panreac Química S.L.U. C/Garraf, 2 Polígono Pla de la Bruguera E-08211 Castellar del Vallès (Barcelona) España Tel. (+34) 937 489 400 Fax (+34) 937 489 401 e-mail: info.es@itwreagents.com

www.itwreagents.com

Till

Quality Assurance Director Director de Garantía Calidad

P. Verge

**CODE: 181693 BATCH:** 0001365644