

**ANALYSIS CERTIFICATE**

**CODE: 251176**

**BATCH: 0000942884**

**PRODUCT: Toluidine Blue O (C.I. 52040) for clinical diagnosis**

**ISSUE DATE: 20/09/2016**

**RETEST DATE: 09/2022**

<b>SPECIFICATIONS</b>	<b>GUARANTEE VALUE</b>	<b>ACTUAL VALUE</b>
Identity	IR passes test	IR passes test
Ratio lmax. P +/- 15 nm	1,00 - 1,23	1,09
A 1% max.	>750	1408
l of max. ABS in H2O	630 - 635 nm	631 nm
T.L.C.	passes test	Passes test
Maximum limit of impurities		
Loss on drying at 135°C	10%	3%

**Panreac Química S.L.U.**  
C/Garraf, 2  
Polígono Pla de la Bruguera  
E-08211 Castellar del Vallès  
(Barcelona) España  
Tel. (+34) 937 489 400  
Fax (+34) 937 489 401  
e-mail: info.es@itwreagents.com  
[www.itwreagents.com](http://www.itwreagents.com)



Quality Assurance Director  
Director de Garantía Calidad  
**P. Verge**

**CODE: 251176**

**BATCH: 0000942884**