

ANALYSIS CERTIFICATE

CODE: 251176

BATCH: 0001228967

PRODUCT: Toluidine Blue O (C.I. 52040) for clinical diagnosis

ISSUE DATE: 28/09/2017

RETEST DATE: 09/2023

SPECIFICATIONS	GUARANTEE VALUE	ACTUAL VALUE
Identity	IR passes test	IR passes test
Ratio lmax. P +/- 15 nm	1,00 - 1,23	1,09
A 1% max.	>750	1408
l of max. ABS in H2O	630 - 635 nm	631 nm
T.L.C.	passes test	Passes test
Maximum limit of impurities		
Loss on drying at 135°C	10%	3%

Panreac Química S.L.U.
C/Garraf, 2
Polígono Pla de la Bruguera
E-08211 Castellar del Vallès
(Barcelona) España
Tel. (+34) 937 489 400
Fax (+34) 937 489 401
e-mail: info.es@itwreagents.com
www.itwreagents.com



Quality Assurance Director
Director de Garantía Calidad
P. Verge

CODE: 251176

BATCH: 0001228967