

ANALYSIS CERTIFICATE

CODE: 251332

BATCH: 0001039352

PRODUCT: Fuchsin Basic (C.I. 42510) for clinical diagnosis

ISSUE DATE: 09/02/2017

RETEST DATE: 02/2023

SPECIFICATIONS	GUARANTEE VALUE	ACTUAL VALUE
Identity	IR passes test	IR passes test
Ratio lmax. P +/- 15 nm	1,16 - 1,35	1,18
A 1% max.	>1600	2504
l of max. ABS in C ₂ H ₅ OH 50%	549 - 552 nm	552 nm
T.L.C.	passes test	Passes test
Maximum limit of impurities		
Loss on drying at 135°C	15%	2%

Panreac Química S.L.U.
C/Garraf, 2
Polígono Pla de la Bruguera
E-08211 Castellar del Vallès
(Barcelona) España
Tel. (+34) 937 489 400
Fax (+34) 937 489 401
e-mail: info.es@itwreagents.com
www.itwreagents.com



Quality Assurance Director
Director de Garantía Calidad
P. Verge

CODE: 251332

BATCH: 0001039352