

**ANALYSIS CERTIFICATE**

**CODE:** 251758

**BATCH:** 0001300579

**PRODUCT:** **Brilliant Green (C.I. 42040) for clinical diagnosis**

**ISSUE DATE:** 08/02/2018

**RETEST DATE:** 02/2024

<b>SPECIFICATIONS</b>	<b>GUARANTEE VALUE</b>	<b>ACTUAL VALUE</b>
Identity	IR passes test	IR passes test
Ratio lmax. P +/- 15 nm	0,94 - 1,14	1,03
A 1% max.	>1750	2243
l of max. ABS in C2H5OH 50%	628 - 632 nm	629 nm
T.L.C.	passes test	Passes test
Maximum limit of impurities		
Loss on drying at 135°C	5%	3%

**Panreac Química S.L.U.**  
C/Garraf, 2  
Polígono Pla de la Bruguera  
E-08211 Castellar del Vallès  
(Barcelona) España  
Tel. (+34) 937 489 400  
Fax (+34) 937 489 401  
e-mail: info.es@itwreagents.com  
[www.itwreagents.com](http://www.itwreagents.com)



Quality Assurance Director  
Director de Garantía Calidad  
**P. Verge**

**CODE:** 251758

**BATCH:** 0001300579